

04-10-2023 12:55:11  
 10.4680724  
 77.6691818



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Chatrapatti Details About Robust Analysis

Lab No: 12162 Detail About Monthly Summary

Month: July Detail about SDI

Year: 2023

Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Date of Result Entered : 14/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	110.3 mg/dL	0.16	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	121 mg/dL	1.72	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by  
**Swiftware Solutions**

**MEDICAL OFFICER**  
 GOVT.PRIMARY HEALTH CENTRE,  
 CHATRAPATTI-624 614.