

04-10-2023 12:53:41

10.4673899

77.6735334



# CHRISTIAN MEDICAL COLLEGE VELLORE

*In Service of the Nation Since 1900...*

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Chatrapatti	<a href="#">Details About Robust Analysis</a>
Lab No	12162	<a href="#">Detail About Monthly Summary</a>
Month	March	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023


Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	72.8 mg/dL	-2.74	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.77 mg/dL	3.67	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	105.6 mg/dL	-0.92	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Site Maintained by



*Goverman*  
**MEDICAL OFFICER**  
GOVT. PRIMARY HEALTH CENTRE,  
CHATRAPATTI - 624 614.