

04-10-2023 12:54:07  
 10.4673907  
 77.6735328



**CHRISTIAN MEDICAL COLLEGE**  
**VELLORE**  
*In Service of the Nation Since 1900...*

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Chatrapatti Details About Robust Analysis  
 Lab No: 12162 Detail About Monthly Summary  
 Month: April Detail about SDI  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	120 mg/dL	0.08	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.7 mg/dL	3	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	117 mg/dL	0.93	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by  
**Swiftware Solutions**

*Copy*  
**MEDICAL OFFICER**  
 GOVT. PRIMARY HEALTH CENTRE,  
 TEL: 024 614.