

VIEW LAB MONTHLY SUMMARY

| | | |
|-------------------|---------------------------|---|
| Lab Name | Govt. PHC, Thengumaragada | Details About Robust Analysis |
| Lab No | 12473 | Detail About Monthly Summary |
| Month | July | Detail about SDI |
| Year | 2023 | |
| Constituent Group | Chemistry III | |

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 17/08/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | Participants | | | Your Value | SDI | SI |
|-------|----------------|---------------------------------------|--|--------------------|--------------|------|------|-------------|-------|------|
| | | | | | DV | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1876 | 109.35 | 5.55 | 6.07 | 97.55 mg/dL | -1.94 | 0.25 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1850 | 5.83 | 6.00 | 0.35 | 6.08 mg/dL | 0.74 | 0.02 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1776 | 113.00 | 4.12 | 4.66 | 109.7 mg/dL | -1.74 | 0.22 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |