

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Manjini Lab No 12574

Constituent Group Chemistry III Date of Result Entered : 20/07/2023

PT item Lyophilized human serum based Date of Report Published : 07/08/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11
SI.IVC						CV	SD	Value	301	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	118 mg/dL	1.43	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.3 mg/dL	-1.51	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	107 mg/dL	-1.29	0.22

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. PHC, Manjini Manjini

Manjini

TAMILNADU636141

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******