



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023

Lab Name: Govt PHC, Armapattinam
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No:
Date of Result Entered:
Date of Report Published:

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Partic
						CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23

SDI Range	Interpretation
Within ± 1.00 to ± 1.00	Excellent
Within ± 1.01 to ± 2.00	Good
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal
Beyond ± 3.0	Unacceptable performance. Action Signal

LAB ADDRESS :
Govt PHC, Armapattinam
Armapattinam
Armapattinam
TAMILNADU 614617

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christal
Dr. Pamela Christal
CMC EQAS Coordin
Christian Medical College

Homogeneity and Stability of the sample is passed
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

M. Nandren Fathima
MEDICAL OFFICER
GOVT PRIMARY HEALTH C
AMMAPATTINAM-614617

