



"Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :		CERTIFICATE NO	
GOVERNMENT PRIMARY HEALTH CENTRE KALLOORANI, VIRUDHUNAGAR, TAMIL NADU-626105		TICS/2023/KAL/VNR/09	
CALIBRATION DATE		EQUIPMENT NAME	
13/01/2023		MICROPIPETTE	
CALIBRATION DUE DATE		MANUFACTURE MODEL	
12/01/2024		THERMOSCIENTIFIC	
		SERIAL NO	
		100-1000 µL	
		INSTALLED DEPARTMENT	
		RW02979 LAB	

1. Have more dust in machine outside Yes No
2. Any Physical damaged in machine Yes No
3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPI22/07/884-01	27-JUL-2023

TEST REPORT					
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.1	+ 1	PASS
		1000	999.3	+ 1	

Temperature 25°C
Humidity RH 55%

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.


ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
Authorized Sign With Seal

Calibrated By: *Shreef*



[Signature]

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
 ☎ 04633-210998 ☑ +91 72000 04781, +91 95001 13759
 ✉ transindiacaibration@gmail.com 🌐 www.transindiacaibration.com
 CIN: U52100TN2020PTC139842





TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED
"Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS:

GOVERNMENT PRIMARY HEALTH CENTRE
KALLOORANI, VIRUDHUNAGAR,
TAMIL NADU-626105

CERTIFICATE NO

TICS/2023/KALL/VNR/08

EQUIPMENT NAME

MICROPIPETTE

MANUFACTURE MODEL

THE RMOSCENTIFIC

SERIAL NO

10-100 µL

INSTALLED DEPARTMENT

RW01342
LAB

CALIBRATION DATE

13/01/2023

CALIBRATION DUE DATE

12/01/2024

1. Have more dust in machine outside
 Yes No
2. Any Physical damaged in machine
 Yes No
3. Machine inside Cleaning (if required)
 Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPL22/07/884-01	27-JUL-2023


TEST REPORT					
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.4	± 1	PASS
		99	99.3	± 1	

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Calibrated By: _____ Authorized Sign With Seal _____



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
 ☎ 04633-210998 📠 +91 72000 04781, +91 95001 13759
 ✉ transindiocalibration@gmail.com 🌐 www.transindiocalibration.com
 CIN: U52100TN2020PTC139842





TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED

"Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :

GOVERNMENT PRIMARY HEALTH CENTRE
KALLOORANI, VIRUDHUNAGAR,
TAMIL NADU-626105

CALIBRATION DATE	13/01/2023	CERTIFICATE NO	TICS/2023/KALLANR/06
CALIBRATION DUE DATE	12/01/2024	EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE MODEL	MICROLUX
		SERIAL NO	5-50 µL
		INSTALLED DEPARTMENT	KALL-06 LAB

- Have more dust in machine outside
 Yes No
- Any Physical damaged in machine
 Yes No
- Machine inside Cleaning (if required)
 Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPI22/07/884-01	27-JUL-2023

TEST REPORT					
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.4	+1	PASS
		50	49.3	+1	PASS

Temperature 25°C
Humidity RH 55%

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By: *Shreef* Authorized Sign With Seal *[Signature]*



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759
✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com
CIN: U52100TN2020PTC139842



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :		CERTIFICATE NO	
GOVERNMENT PRIMARY HEALTH CENTRE KALLOORANI, VIRUDHUNAGAR, TAMIL NADU -626105		TICS2023/KALL/VNR/01	
MANUFACTURE NAME		REFRIGERATOR	
SAMSUNG		SAMSUNG	
MODEL		NA	
13/01/2023		KALL-01	
CALIBRATION DATE		LAB	
12/01/2024		LAB	
DATE		DEPARTMENT	

- General maintenance required Yes No
- Any Physical damage observed Yes No
- Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DATA LOGGER WITH SENSOR	TEMPSONS	TRUB-96	2022-2023/esp/2205005-01

TEST REPORT

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
TEMPERATURE	°C	5	5.2	± 1	PASS

* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :  Authorized Sign With Seal 



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 62780
 ☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759
 ✉ transindiocalibration@gmail.com @ www.transindiocalibration.com
 CIN: U52100TN2020PTC139842



TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED
 "Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :

GOVERNMENT PRIMARY HEALTH CENTRE
 KALLOOKANI, VIRUDHINAGAR,
 TAMIL NADU-626105

CALIBRATION DATE	13/01/2023	CERTIFICATE NO	TICS/SD023/KALLVNR/05
CALIBRATION DUE DATE	12/01/2024	EQUIPMENT NAME	MICROSCOPE
		MANUFACTURE MODEL	BIJSCO
		SERIAL NO	BI.S116
		INSTALLED DEPARTMENT	8712 LAB

- General maintenance required
 Yes No
- Any Physical damage observed
 Yes No
- Spare Replacement (if required)
 Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFETY ANALYZER	DATREND	Vpad-miri	VPM 18040011

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Calibrated By: _____ Authorized Sign With Seal

Shree



BR

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
 ☎ 04633-210998 📠 +91 72000 04781, +91 95001 13759
 ✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com
 CIN: U52100TN2020PTC139842



TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED

"Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS:
GOVERNMENT PRIMARY HEALTH CENTRE
KALLOORANI, VIRUDHUNAGAR,
TAMIL NADU-626105

CALIBRATION DATE	13/01/2023	CERTIFICATE NO	TIC/S/2023/KALL/VNR/07
CALIBRATION DUE DATE	12/01/2024	EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE MODEL	MICROLUX
		SERIAL NO	100-1000 µL
		INSTALLED DEPARTMENT	KALL-07 LAB

1. Have more dust in machine outside Yes No
2. Any Physical damaged in machine Yes No
3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSPI22/07/884-01	27-JUL-2023

TEST REPORT					
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.3	±1	PASS
		1000	999.2	±1	PASS

* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By: _____ Authorized Sign With Seal



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
 ☎ 04633-210998 ☑ +91 72000 04781, +91 95001 13759
 ✉ transindiacalibration@gmail.com @ www.transindiacalibration.com
 CIN: U52100TN2020PTC139842



CUSTOMER DETAILS :
 GOVERNMENT PRIMARY HEALTH CENTRE
 KALLOORANI, VIRUDHUNAGAR,
 TAMIL NADU-626105

CALIBRATION / TESTING REPORT

CALIBRATION DATE	13/01/2023	CERTIFICATE NO	TICS/2023/KAL/VN/R/02
CALIBRATION DUE DATE	12/01/2024	EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	BD-153
		SERIAL NO	76664
		INSTALLED DEPARTMENT	LAB

- General maintenance required
 Yes No
- Any Physical damage observed
 Yes No
- Spare Replacement (if required)
 Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MICO	KM2234 BL	S1049875
TEST REPORT			
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES
ROTATION PER MINUTE	RPM	100	103
		1000	1002
			± 5
			PASS

*The above mentioned instrument has been tested using standards manufactures manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By: *[Signature]*

Authorized Sign With Seal *[Signature]*



No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.
 ☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759
 ✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com
 CIN: U52100TN2020PTC139842



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :
 GOVERNMENT PRIMARY HEALTH CENTRE,
 KALLIOPPANI, VIRUDHINAGAR,
 TAMIL NADU -626105

CALIBRATION DATE	13/01/2023	CERTIFICATE NO	TICS/2023/KALLIANR/03
CALIBRATION DUE DATE	12/01/2024	EQUIPMENT NAME	CELL COUNTER
		MANUFACTURE	SYSMEX
		SERIAL NO	XP100
		INSTALLED	35603
		DEPARTMENT	LABORATORY

1. General maintenance required Yes No
2. Any Physical damage observed Yes No
3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFETY ANALYZER	DATREND	V Pad-mma	VPM18040011

*The above mentioned instrument has been tested using standards manufacturer's recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By:



Authorized Sign With Seal



TRANSINDIA CALIBRATION SERVICES PRIVATE LIMITED
 "Perfecting the art calibration in Medical Equipment's"

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS :

GOVERNMENT PRIMARY HEALTH CENTRE
 KALLOORANI, VIRUDHUNAGAR,
 TAMIL NADU-626105

CERTIFICATE NO	TICS/2023/KAL/VNR/04
EQUIPMENT NAME	SEMI AUTO ANALYZER
MANUFACTURE MODEL	ROBONIK
SERIAL NO	PRITEST
INSTALLED DEPARTMENT	ATCD3200818RBK LAB
CALIBRATION DATE	13/01/2023
CALIBRATION DUE DATE	12/01/2024

- General maintenance required Yes No
- Any Physical damage observed Yes No
- Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFETY ANALYZER	DATREND	V Pad-mini	VPN118040011

* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

Calibrated By: ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
 Authorized Sign With Seal



Blanch

[Signature]

No.763/19, Block No:10, KTC Nagar, Kulassekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

☎ 04633-210998 📠 +91 72000 04781, +91 95001 13759

✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com

CIN: U52100TN2020PTC139842