



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. PHC, Kallankattuvalasu
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 7456
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	170 mg/dL	-0.62	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22
SDI Range Interpretation										
				Within -1.00 to +1.00	Excellent.					
				Within ±1.01 to ±2.00	Good.					
				Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
				Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS :
 Govt. PHC, Kallankattuvalasu
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT.
 Kallankattuvalasu
 TAMILNADU638008

MEDICAL OFFICER
 Govt. Primary Health Centre
 KALLANKATTUVALASU
 Namakkal Dist. 638 183

Coordinator Contact Details:
 Email: cllngc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paola Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential