

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kallankattuvalasu  
 Lab No 7456  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	110 mg/dL	0.11	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links