

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kallankattuvalasu
 Lab No 7456
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	185.9 mg/dL	0.53	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.33	108.7 mg/dL	-1.01	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

[Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre
 KALLANKATTUVALASU
 Namakkal Dist-638 183

Links

- [Modify Address](#)
- [Modify Method](#)
- [Submit Lab Results](#)
- [View Lab Monthly Summary](#)
- [View Summary](#)
- [Change Password](#)
- [Home](#)
- [Sign-out](#)