

**CUSTOMER ADDRESS: GENERAL PRIMARY HEALTH CENTRE,
ELANTHAKOTTAL,
NAMAKKAL DISTRICT**

	DISCIPLINE	: MECHANICAL CALIBRATION
NAME OF EQUIPMENT:	MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO	: NBS210847251	DATE OF ISSUE : 03.10.2023
RECEIPT NO	: 10136	TEMPERATURE : 24.5 ⁰ C
RECEIPT DATE	: 03.10.2023	HUMIDITY : 53%
JOB CARD NO	: TPM000	AIR PRESURE : 959hpa
JOB CARD DATE	: 03.10.2023	MEDIUM OF WEIGHING : GRADE III WATER
		DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE	: 100-1000 μ l	DENSITY CORRECTION
		FACTOR[Z]: : 1.0037 μ l/mg
TIPS USED	: 100-1000 μ l	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE ACCREDITATION: VOLUME-MICROPIPETTE (100μl-1000μl)		

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT : THERMO SCIENTIFIC
S.L NO : RW02690
RESOLUTION : 0.00001g up to 82g & above is 0.0001 g
CALIBRATED BY : NEWTONBIOMEDICAL SERVICE
CALIBRATION VALIDITY : 02.10.2024

