

CALIBRATION CERTIFICATE

DATE: 03.10.2023

NAME OF CUSTOMER:
GENERAL PRIMARY HEALTH CENTER
ELANTHAKOTTAI

DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE

MODEL NO: BD-153

SL.NO: 76698

COMPANY: BDI

INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER

MODEL NO:KUSAM-MECO&KM-2234BL

SL.NO: 76698

CERTIFIED BY: MSIR INDIA PVT.

<i>REGULATOR STEEPS</i>	<i>INSTRUMENT SPEED</i>	<i>TACHOMETER READING</i>	<i>ALLOWABLE ERROR</i>	<i>ERROR OBSERVED ±</i>	<i>RESULT</i>
1.	700	716	±50	16	OK
3.	2100	2060	±50	40	OK
5.	3500	3480	±50	20	OK

NEXT CALIBRATION DUE ON: 02.10.2024

TESTED BY: G DINESH

APPROVED BY: Q.A

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

