

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kallankattuvalasu
 Lab No 7456
 Month April
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data


[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	QV	Cv	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	118.9 mg/dL	-0.08	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.38 mg/dL	1.26	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110.3 mg/dL	-0.23	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


MEDICAL OFFICER
 Govt. Primary Health Centre
 KALLANKATTUVALASU
 Namakkal Dist. 638 183

Links

[Modify Address](#) [Modify Methods](#) [Submit Lab Results](#) [View Lab Monthly Summary](#)
[Forgot Password](#) [Home](#) [Logout](#)