



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. Block PHC, Uppidamangalam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 7409
Date of Result Entered : 21/08/2023
Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Mean	SDI	SI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	128.34 mg/dL	-5.25	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2.41 mg/dL	5.42	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	127.69 mg/dL	3.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. Block PHC, Uppidamangalam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, KARUR DISTRICT.
Uppidamangalam
TAMILNADU 639114

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

[Signature]
DISTRICT MICROBIOLOGIST
DISTRICT PUBLIC HEALTH LAB
GOVT. DISTRICT HQ. HOSPITAL
KULITHALAI (TK)
KARUR. (DT)

[Signature]
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential