



	ELECTRICA	AL SAFETY TESTING/ CALIBRATION		
Report No	: HBMS/CAL/23-24	Calibration Date : 22.09.2023		
		Calibration Valid: 21.03.2024		
1.1CUSTO	OMER DETAILS:			
Name and Address of the organization		M/S, PRIMARY HEALTH CENTRE, KALLANKATTUVALASU,KUMARAPALAYAM,		
		NAMAKKAL.		
1.2. DESC	RIPTION OF DEVICE UNDER TES	ST (DUT):		
EQUIPMI	ENT NAME	SEMI AUTO ANALYZER		
i	Manufactured by	ROBONIK		
ii	Model	PRIETEST TOUCH		
iii	Serial No	ATCD0700621RBK		
iv	Location	LABORATORY		

2.CONDITION OF T	HE ITEM	WHEN	RECEIVED

No visible damages and in working order

3.TEST DONE	Electrical safety/ Performance Testing

4. STATUS						
Manufactures Specification	User specification	Within specification	Out of specification	Calibration	Electrical Safety Test	Performance Analysis
1	-	√	-	-	1	-

T

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

0

99400 28355, 9751206106

88833 15662

A

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com

M	ed	ica	IS	VS	te	m	S
	Designation	نظفتان		, -	-	-	r.

S.no	Name of the instrument	Make	Model	Serial no	Cal Due
	Electrical safety	Fluke	Esa612	3005050	07-OCT-202

s.NO	PARAMETER	OBSERVED VALUE	ACCEPTANCE LIMITS AS PER THE STD	REMARKS
1.	Protective earth resistance	0.129Ω	<0.3 Ω	Ok
2.	Earth leakage	134.21μΑ	<210 µA NC	Ok
3.	Patient leakage current	NA	<100 µA BF	Ok
4.	Equipment current	0.3A	As per the manufacture spec	Ok
5.	Mains voltage (P-E)	230.5V	ma.	Ok
6.	Mains voltage (P-N)	231.2V	-	Ok
7.	Mains voltage (N-E)	1.0V	<5V	Ok
8.	Enclosure leakage current	50 µА	<100 μA (normal)	Ok
9.	Insulation resistance	OL	>100 M Ω	Ok

S.NO	7.Remarks
1.	This report is applicable to the sample tested only
2.	The instrument used for testing are valid calibration and are traceable to national standards
3.	Parameter of the DUT were verified and found to be within the specified limits

Verified by:

MD Reg No: TN/CBE/MD4Z/00016

GSTIN: 33AAJFH1833M1ZB

0

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com