


VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Pilathur	Details About Robust Analysis Detail About Monthly Summary Detail about SDI	
Lab No	10526	<input type="button" value="All Analyser Result"/> <input type="button" value="Print"/>	
Month	May	Date of Result Entered : 20/05/2023	
Year	2023	Date of Report Published : 05/06/2023	
Constituent Group	Chemistry III	Click on the analyte to view Graphical Data	

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	246 mg/dL	6.94	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	86 mg/dL	-4.91	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links


MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
PILATHUR
DINDIGUL DIST.