

Medical Systems

|                                      | ELECTRICA                   | AL SAFETY TESTING/ CALIBRATION |
|--------------------------------------|-----------------------------|--------------------------------|
| Report No                            | : HBMS/CAL/23-24            | Calibration Date : 30.09.2023  |
|                                      |                             | Calibration Valid: 29.03.2024  |
| 1.1CUSTO                             | OMER DETAILS:               |                                |
| Name and Address of the organization |                             | M/S, PRIMARY HEALTH CENTRE,    |
|                                      |                             | SEETHARAMPALAYAM, NAMAKKAL.    |
| 1.2. DESC                            | RIPTION OF DEVICE UNDER TES | ST (DUT):                      |
| EQUIPMI                              | ENT NAME                    | SEMI AUTO ANAL YZER            |
| i                                    | Manufactured by             | ROBONIK                        |
| ii                                   | Model                       | PRIETEST T OUCH                |
| iii                                  | Serial No                   | ATP0550122RBK                  |
| iv                                   | Location                    | LABORATORY                     |

| 2.CONDITION OF THE ITEM WHEN RECEIVED   |              |
|---|--------------|
| NY 4413 4                               | a control of |
| No visible damages and in working order |              |

| 3.TEST DONE | Electrical safety/ Performance Testing |
|-------------|--|
|             | Electrical safety, reflormance resung  |
|             |  |
|             |  |

| 4. STATUS                     |                       |                      |                      |             |                           |                         |
|-------------------------------|-----------------------|----------------------|----------------------|-------------|---------------------------|-------------------------|
| Manufactures<br>Specification | User<br>specification | Within specification | Out of specification | Calibration | Electrical<br>Safety Test | Performance<br>Analysis |
| √                             | -                     | <b>√</b>             | -                    | -           | 1                         | -                       |

0

0

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



Medical Systems

st & Care

| S.TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING |                               |       |        |           |             |
|--|-------------------------------|-------|--------|-----------|-------------|
| S.no   | Name of the instrument        | Make  | Model  | Serial no | Cal Due     |
| 1.   | Electrical safety<br>analyser | Fluke | Esa612 | 3005050   | 07-OCT-2023 |

| ELEC | TRICAL SAFETY TEST          |                |                                     |         |
|------|-----------------------------|----------------|-------------------------------------|---------|
| s.NO | PARAMETER                   | OBSERVED VALUE | ACCEPTANCE LIMITS AS<br>PER THE STD | REMARKS |
| 1.   | Protective earth resistance | 0.129Q         | <0.3 Ω                              | Ok      |
| 2.   | Earth leakage               | 134.21μΑ       | <210 μA NC                          | Ok      |
| 3.   | Patient leakage current     | NA             | <100 µA BF                          | Ok      |
| 4.   | Equipment current           | 0,3A           | As per the manufacture spec         | Ok      |
| 5.   | Mains voltage (P-E)         | 230.5V         | <u> </u>                            | Ok      |
| 6.   | Mains voltage (P-N)         | 231.2V         | <del>-</del>                        | Ok      |
| 7.   | Mains voltage (N-E)         | 1.0V           | <5V                                 | Ok      |
| 8.   | Enclosure leakage current   | 50 μA          | <100 μA (normal)                    | Ok      |
| 9.   | Insulation resistance       | OL             | >100 M Q                            | Ok      |

| S.NO | 7.Remarks   |
|------|---|
| 1.   | This report is applicable to the sample tested only   |
| 2.   | The instrument used for testing are valid calibration and are traceable to national standards |
| 3.   | Parameter of the DUT were verified and found to be within the specified limits                |

Verified by :

MD Reg No: 14 200 E/MD42/00016

GSTIN: 33AAJFH1833M1ZB

0

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE-04, hbmedical system@gmail.com