

	CAL	IBRA	TION	CERT	IFICAT	<u>E</u>	
CERTIFICATE NO:	SBS/CL/23/10363			ME	CHANICAL	Pa	age No:1 of 1
Issue Date			18-08-20	23			
SRF No & Date			SRF/23/	00320-0002	& 17-08-2023		
Receipt Date			17-08-20	23			
Calibration Date			17-08-20	23			
Calibration Due			16-08-2024				
Customer Name & Add	ress						
GOVERNMENT COMM	UNITY HEALTH CENTR	RE,					
PUDUPETTAI- 635651,	TIRUPATTUR DISTRIC	CT.					
	D	etails of D	evice Un	der Calibra	tion (DUC)		
Description :	CENTRIFUGE		Make & Model		;	BD INSTRUMENTATION & BD-153	
Range :	5000 RPM		Sr. No		:	76659	
Resolution :	50 RPM		Identification No		:	NA	
DUC Condition :	Satisfactory		Location		:	LABORATORY	
	Environmenta	I Condition	ns & Stan	dard Opera	ting Procedu	re Details	
Environmental Details Temperature		ure: 24.3 °	e: 24.3 °C R		midity	54 % Rh	
Calibration Procedure No SBS/CP/ML/		/L/04	04 Calib		done at	ONSITE	
		Refe	rence Sta	ndards Det	ails		
S.No Description	Make/ S	Make/ SI No:		Certificate No		Validity	
1 Digital Tachometer			Lutron & Q638380		JRPM-CCTR-A&S-2022-0036		03-10-2023
CALIBRATION RESULT	ſS						

DEVICE UNDER S.No STANDARD INSTRUMENTS DEVIATION EXPANDED UNCERTAINTY (±) CALIBRATION READINGS RPM RPM RPM % 1000 999.5 0.5 4.2 1 2 2000 1999.7 0.3 4.2 3000 2999.5 3 0.5 4.2

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.


4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) M.RAGUL



Quality Manager (D.VETRI SELVI) Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.