

VIEW LAB MONTHLY SUMMARY

| | | |
|-------------------|------------------------|-------------------------------|
| Lab Name | Govt. PHC, Arasampattu | Details About Robust Analysis |
| Lab No | 13121 | Detail About Monthly Summary |
| Month | April | Detail about SDI |
| Year | 2023 | |
| Constituent Group | Chemistry III | |

Click on the analyte to view Graphical Data

Date of Result Entered : 20/04/2023


Date of Report Published : 06/05/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|-------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1832 | 119.45 | 6.04 | 7.21 | 119.1 mg/dL | -0.05 | 0.34 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1820 | 2.16 | 8.09 | 0.18 | 2 mg/dL | -0.91 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Any Analyser (Automation / Semi Automation) | 143 | 111.22 | 4.31 | 4.79 | 114 mg/dL | 0.58 | 0.80 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |


MEDICAL OFFICER,
GOVT.PRIMARY HEALTH CENTRE
ARASAMPATTU-606 401