



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name Govt. PHC, Arasampattu Lab No 13121
Constituent Group Chemistry III Date of Result Entered : 21/02/2023
PT item Lyophilized human serum based Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	165 mg/dL	0.97	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.7 mg/dL	0.98	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	112.46	5.14	5.78	108 mg/dL	-0.77	0.97

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Arasampattu
Arasampattu
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

S. Shaamil
MEDICAL OFFICER,
GOVT. PRIMARY HEALTH CENTRE
ARASAMPATTU-606 401.