



heartbeat
Medical Systems

S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :




MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navalndia Road, Peelamedu,
CBE - 04, hbmedicalsystem@gmail.com



heartbeat
Medical Systems

CALIBRATION REPORT

Report No : HBMS-CAL/23-24		Calibration Date : 04.09.2023
		Calibration Valid: 03.03.2024
I.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, MANUR, THIRUNELVELL
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BD INSTRUMENTATION
iii	Serial No	BDI-152
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For - Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1003	3
2	1200	1198	-2
3	1500	1499	-1
4	2000	2000	0

Acceptable deviation is \pm 5 RPM if depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navalndia Road, Peelam

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail