



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. UGPHC, Erwadi Lab No: 7478
 Constituent Group: Chemistry III Date of Result Entered: 20/07/2023
 PT item: Lyophilized human serum based Date of Report Published: 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	106 mg/dL	-0.55	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. UGPHC, Erwadi
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.
 Erwadi
 TAMILNADU623608

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Dr. Pamela Christudoss
 மருத்துவ அலுவலர்
 சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி.
 சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி.
 சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி.
 சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி. சி.ம.சி.