



# CHRISTUMAN MEDICAL COLLEGE

VELLORE

*In Service of the Nation Since 1900...*



# CHRISTUMAN

*In Service of the Nation Since 1900...*

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Bargur  
 Lab No: 42177  
 Month: July  
 Year: 2023  
 Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	120 mg/dL	1.76	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.45 mg/dL	-1.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	113 mg/dL	1.07	0.22
SDI Range Interpretation										
Within -1.00 to +1.00 Excellent.										
Within ±1.01 to ±2.00 Good.										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.										
Beyond ±3.0 Unacceptable performance. Action Signal.										

Links	
Modify Address	Modify Methods
Yearly Summary	Change Password
	Submit Lab Results
	View Lab Monthly Summary
	Sign-out
	Home