



NEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Bargur  
Lab No: 12177  
Month: June  
Year: 2023  
Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	307.5 mg/dL	1.41	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	115.9 mg/dL	0.43	0.24
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

Modify Address	Modify Methods	Submit Lab Results	New Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out