



CHRISTIAN MEDICAL COLLEGE VELLORE

In Service of the Nation Since 1900...



CHRISTIAN

In Service of

NEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Bangur

Lab No: 12177

Month: March

Year: 2023

Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

[Click on the analyte to view Graphical Data](#)

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			SDI	U
						CV	SD	Your Value		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	96 mg/dL	0.66	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.2 mg/dL	-1.07	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110.2 mg/dL	-0.13	0.28
SDI Range										
Interpretation										
Within -1.00 to +1.00										
Excellent.										
Within ±1.01 to ±2.00										
Good										
Within ±2.01 to ±2.99										
Accept with caution. Warning Signal.										
Beyond ±3.0										
Unacceptable performance. Action Signal.										

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

[Change Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)