



G MULTI SERVICES

78, Yahooob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452.
gmultiservicesgobi@gmail.com. Cell : 96597 20654.
Bio Medical Instruments Services

Er. A. Ibathul Razip DECE.,BE (ICE)
Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment: **CENTRIFUGE** PROTECT CLASS/TEST : Risk class I
MAKE : C-85414 S.No : ZDCN-07510
MODEL : Remi Locatio. : GPHC, Bargur

S.no	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range :210-240v) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage current (f) Battery conditions & charging status	235 volts 0 to 5volts No Zero ohms 10mA NA	Stabilizer/Ups With in Ranges Well
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets,motors assembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked Ok
3	Performance Test (a) ON /OFF switch (b) Speed regulator switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal Ok Accurate Ok No Well Well	Ok Perfect Perfect
4	Calibration Test (a) Vibration test (b) Speed regulation gradual hike/wobbling test (c) Running test with load (d) Display checkup (e) Timer test	Done Ok	Perfect
Formulated By: Edition Research Laboratory		Issued By:	User Department:

Endorsed By: Research Team Head			
Reviewed On: 26.03.23 Next Review : 25.03.24			