

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Chinnapalli Kuppam	Details About Robust Analysis
Lab No	13053	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/06/2023


Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	296.6 mg/dL	0.67	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.49 mg/dL	-0.47	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	111.13 mg/dL	-0.51	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE,
 CHINNAPALLI