



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt. UGPHC, Erwadi  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 7478  
Date of Result Entered: 20/03/2023  
Date of Report Published: 04/04/2023

Sl. No.	Analyte	Reference / Participating Item	Analytical Method	No. of Participants	SD	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	93.96 mg/dL	0.36	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.37 mg/dL	0.33	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	108.47 mg/dL	-0.42	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. UGPHC, Erwadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, RAMANATHAPURAM DISTRICT.  
Erwadi  
MILNADU623608

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*A.G.*