

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Chinnappalli Kupparam
 Lab No 13053
 Month August
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

| Sl.No | Analyte | Method / Principle Name | Analyser Name | No of Participants | DV | Participants CV | SD | Your Value | SDI | U |
|--------------------------|----------------|---------------------------------------|--|--|--------|-----------------|------|--------------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 180.43 mg/dL | 0.54 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.8 mg/dL | 0.34 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 111.76 mg/dL | -0.44 | 0.22 |
| SDI Range Interpretation | | | | | | | | | | |
| Within -1.00 to +1.00 | | | | Excellent. | | | | | | |
| Within ±1.01 to ±2.00 | | | | Good. | | | | | | |
| Within ±2.01 to ±2.99 | | | | Accept with caution. Warning Signal. | | | | | | |
| Beyond ±3.0 | | | | Unacceptable performance. Action Signal. | | | | | | |

Links