

6/14/23, 3:03 PM

External Quality Assurance Scheme - Print Monthly Summary



CHRISTIAN MEDICAL COLLEGE  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PHC, Avoor  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 10569  
Date of Result Entered: 19/05/2023  
Date of Report Published: 05/06/2023

| SL No | Analyte        | Method / Principle Name               | Analyser Name                                | No of Participants | DV     | CV   | Participants | Your Value | SD   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|------|--------------|------------|------|------|
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1834               | 180.88 | 5.19 | 9.39         | 196 mg/dL  | 1.81 | 0.44 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1854               | 1.76   | 7.48 | 0.13         | 1.9 mg/dL  | 1.06 | 0.51 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1798               | 114.60 | 5.08 | 5.83         | 115 mg/dL  | 6.87 | 0.27 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution, Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance, Action Signal. |

LAB ADDRESS :  
Govt. PHC, Avoor  
CLINICAL LABORATORY, PUDUKOTTAI(DIST)  
Avoor  
TAMILNADU622515

Coordinator Contact Details:  
Email:clinqc@cmcvc60ra.ac.in  
Contact Number: 0416-2283102

*Pamela Christodora*  
Dr. Pamela Christodora  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

**MEDICAL OFFICER**  
GOVT. PRIMARY HEALTH CENTRE  
AVOOR-622 515  
PUDUKOTTAI DISTRICT