



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

| | | | |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name | Govt. Block PHC, Periapalayam | Lab No | 7639 |
| Constituent Group | Chemistry III | Date of Result Entered : | 19/08/2023 |
| PT Item | Lyophilized human serum based | Date of Report Published : | 01/09/2023 |

| Sl No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 178 mg/dL | 0.27 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.65 mg/dL | -0.92 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 109 mg/dL | -1.04 | 0.22 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. Block PHC, Periapalayam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVALLUR DISTRICT.
Periapalayam
TAMILNADU 601102

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

