



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. UPHC, Sathuvachari Lab No 9299
 Constituent Group Chemistry III Date of Result Entered : 19/07/2023
 PT item Lyophilized human serum based Date of Report Published : 07/08/2023

S.N.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	110 mg/dl	0.11	0.28
2	CREATININE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5mg	-0.94	0.02
3	CHOLETEROL III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dl	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Sathuvachari
 39TH STREET, PHASE II SATHUVACHARI
 39Th Street, Phase II, Sathuvachari
 TAMILNADU 632009

Pamela Christudoss

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMCEQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 *****End of Report*****

[Signature]
 MEDICAL OFFICER
 Urban Govt. Primary Health Centre
 Sathuvachari - 632009.