



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. UPHC, Sathuvachari Lab No 9299  
 Constituent Group Chemistry III Date of Result Entered : 08/04/2023  
 PT item Lyophilized human serum based Date of Report Published : 06/05/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	125 mg/dl	0.77	0.34
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.31 mg/dl	0.86	0.01
3	CHOLETEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	101.3 mg/dl	-1.78	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### LAB ADDRESS :

Govt. UPHC, Sathuvachari  
 39TH STREET, PHASE II SATHUVACHARI  
 39Th Street, Phase II, Sathuvachari  
 TAMILNADU 632009

*Pamela Christudoss*

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMCEQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not subcontract any components  
 \*\*\*\*\*End of Report\*\*\*\*\*

*[Signature]*  
 MEDICAL OFFICER  
 Urban Govt. Primary Health Centre  
 Sathuvachari - 632009.