

## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12346	MEDICAL DEVICES	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00513-0004 & 29-09-2023	
Receipt Date	29-09-2023	
Calibration Date	29-09-2023	
Calibration Due	29-09-2024	

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE,  
 AVALUR-631605, KANCHIPURAM DISTRICT.

Details of Device Under Calibration (DUC)			
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD0071220RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY


RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	79	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	209	7.3
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	229	7.2

**REMARKS**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
 (Calibration Engineer)  
 E.ESWAR



  
 Technical Manager  
 C.SHANMUGARAJ

Chief Executive