

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Silamarathupatti	<a href="#">Details About Robust Analysis</a>									
Lab No	12701	<a href="#">Detail About Monthly Summary</a>									
Month	July	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 20/07/2023									
		Date of Report Published : 07/08/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Any Analyser (Automation / Semi Automation )	75	109.71	5.43	5.95	120.4 mg/dL	1.80	1.37	
2	<a href="#">UREA III</a>	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation )	11	50.80	9.44	4.79	48.6 mg/dL	-0.46	2.89	
3	<a href="#">CREATININE III</a>	Jaffes Kinetic Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	66	5.89	4.60	0.27	5.89 mg/dL	0.00	0.07	
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	30	5.35	9.89	0.53	5.78 mg/dL	0.81	0.19	
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	112.78	4.33	4.88	112.8 mg/dL	0.00	0.80	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>