

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Silamarathupatti	<a href="#">Details About Robust Analysis</a>									
Lab No	12701	<a href="#">Detail About Monthly Summary</a>									
Month	March	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 20/03/2023									
		Date of Report Published : 04/04/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Any Analyser (Automation / Semi Automation )	74	92.52	7.71	7.13	97.3 mg/dL	0.67	1.66	
2	<a href="#">UREA III</a>	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation )	12	55.01	12.13	6.67	60.5 mg/dL	0.82	3.85	
3	<a href="#">CREATININE III</a>	Jaffes Kinetic Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	70	1.32	7.52	0.10	1.3 mg/dL	-0.20	0.02	
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	43	4.73	7.17	0.34	4.65 mg/dL	-0.24	0.10	
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	111.55	4.29	4.78	108.8 mg/dL	-0.57	0.79	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>