

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Bangalore Road, Konavattam**  
 Lab No **9300**  
 Month **May**  
 Year **2023**  
 Constituent Group **Chemistry III**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)


Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Mean Value	SDI	SI
						CV	SDI			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	198 mg/dL	1.82	0.44
2	CREATININE III	Jaffes Kinotic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	121 mg/dL	1.10	0.27
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within +1.01 to +2.00		Good.								
Within +2.01 to +2.99		Accept with caution. Warning Signal.								
Beyond +3.0		Unacceptable performance. Action Signal.								

Links

<a href="#">Home</a>	<a href="#">Modify Methods</a>	<a href="#">Summit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Home</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
**MEDICAL OFFICER**  
 Urban Primary Health Centre  
 Konavattam, Vellore Corporation.