

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Bangalore Road, Konavattam	Details About Robust Analysis
Lab No	9300	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	256 mg/dL	-3.66	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	103 mg/dL	1017.10	10.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	1 mg/dL	-26.07	0.20

SDI Range

Interpretation

**MEDICAL OFFICER**  
 Urban Primary Health Centre  
 Konavattam, Vellore Corporation.