



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.08.2023
	Calibration valid : 22.02.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, POOLAMPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i   Manufactured by	THERMOSCIENTIFIC
ii   Model	FINNEPIPETTE F3
iii   Value	10 µl -100 µl
iv   Serial No	QW13599

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com

CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 23.08.2023
	Calibration Valid: 22.02.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, POOLAMPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROLUX
II Model	5µl - 50µl
III Serial No	PHC/MP/PP/002
IV Location	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.08.2023 Calibration valid: 22.02.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, POOLAMPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	MICROLUX
ii Model	---
iii Value	100µl-1000µl
iv Serial No	PHC/MP/PP/001

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.05	1000.15
2	100.03	500.04	1000.13
3	100.06	500.03	1000.11
4	100.08	500.01	1000.07
5	100.01	499.99	1000.08
6	100.09	500.09	1000.11

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.08.2023
	Calibration valid: 22.02.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, POOLAMPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNEPIPETTE F3
iii Value	100µl-1000µl
iv Serial No	QW14220

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com