



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/11514

Page No: 1 of 1

Customer Name & Address

GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE,  
ODUGATHUR-632103.

SRF No.	SRI/23/00360-0004
Date	15-09-2023
Receipt Date	15-09-2023
Date of Calibration	15-09-2023
Due Date for Calibration	14-09-2024
Issue Date	16-09-2023

Details of Unit Under Calibration

Description	MICROPIPETTE	Make	ERBA
Range	10µl	Model	BIOHIT
Resolution	FIXED	Material	PVC
Serial Number	7066823	Operating Range	10µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.9	Humidity (%RH)	57	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SL No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.94	9.93	9.95	9.92	9.96	9.94	-0.06	0.02	0.47
		9.95	9.93	9.96	9.97	9.93				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

*C. Sivabalan*  
(Calibration Engineer)  
C.SIVABALAN



Authorised by:

*C. Sivabalan*  
(Quality Manager/Chief Executive)  
C.SIVABALAN

*S. Sridhar*  
BLOCK MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
ODUGATHUR - 632 103.

*Kal*  
மருத்துவ அலுவலர்  
ஆரம்ப சுகாதார நிலையம்  
ஒடுகத்தூர்-632 103.