



VIEW LAB MONTHLY SUMMARY

11081 2023 AUGUST MONTH

Lab Name: Govt. PHC, Esanatham
 Lab No: 10544
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Year Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	130 mg/dl
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2.3 mg/dl
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	12 mg/dl

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Home	Modify Methods	Submit Lab Results	View Lab Mon
Monthly Summary	Change Password	Home	Sign-out

Site Maintained by

Swiftware Solutions

Best Viewed