



Trust & Care



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 13.09.2023
		Calibration Valid: 12.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALRH CENTRE, MELAPALAYAM, THIRUNELVELL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BD INSTRUMENTATION
iii	Serial No	BDI-154
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	999	-1
2	1200	1209	9
3	1500	1508	8
4	2000	2001	1

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :


