

血液细胞分析仪出厂校准原始记录单

Factory Calibration Record of Hematology Analyzers

一、校准基本信息

A. Basic Calibration Information

仪器型号 Instrument Model	BC-20s	序列号S/N Serial No.	TK-24011674
名称 Name	校准物 Calibrator	批号 Reagent Lot	BC10122C-1

二、出厂校准系数

B. Factory Calibration Factors

参数 Parameter	数值 (Numerical Value)
WBC全血校准系数 Calibration Factor(WB Mode)	102.81%
RBC全血校准系数 Calibration Factor(WB Mode)	95.84%
HGB全血校准系数 Calibration Factor(WB Mode)	96.22%
MCV全血校准系数 Calibration Factor(WB Mode)	101.17%
PLT全血校准系数 Calibration Factor(WB Mode)	96.90%
WBC预稀释校准系数 Calibration Factor(PD Mode)	101.71%
RBC预稀释校准系数 Calibration Factor(PD Mode)	91.54%
HGB预稀释校准系数 Calibration Factor(PD Mode)	92.74%
MCV预稀释校准系数 Calibration Factor(PD Mode)	102.43%
PLT预稀释校准系数 Calibration Factor(PD Mode)	93.36%

签字Signature/盖章Seal : 1- 沈 飞

日期/Date : 2022-04-13

