

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Malaikudipatti									
Lab No	12508	Back to Search Page								
Month	July									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 19/07/2023										
Date of Report Published : 07/08/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10			1.70
			Transasia / Erba	24	109.83	6.83	7.50			3.06
			Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	101 mg/dL	-1.38	0.28
			Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95			1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34			0.12
			Transasia / Erba	18	5.79	3.00	0.17			0.08
			Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.5 mg/dL	1.91	0.02
			Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27			0.07
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04			1.39
			Transasia / Erba	19	112.64	4.38	4.93			2.26
			Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	106 mg/dL	-1.50	0.22
			Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88			0.80
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

