

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Malaikudipatti									
Lab No	12508	Back to Search Page								
Month	April									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 20/04/2023										
Date of Report Published : 06/05/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67			2.93
			Transasia / Erba	19	118.42	8.48	10.04			4.61
			Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121 mg/dL	0.21	0.34
			Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83			1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15			0.05
			Transasia / Erba	19	2.14	6.03	0.13			0.06
			Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
			Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03			2.07
			Transasia / Erba	20	111.15	4.25	4.72			2.11
			Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	103 mg/dL	-1.49	0.28
			Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79			0.80
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

