

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Malaikudipatti									
Lab No	12508	Back to Search Page								
Month	March									
Year	2023	Print								
Constituent Group	Chemistry III									
Date of Result Entered : 20/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84			2.61
			Transasia / Erba	18	92.24	11.77	10.85			5.12
			Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	118 mg/dL	3.89	0.31
			Any Analyser (Automation / Semi Automation)	74	92.52	7.71	7.13			1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13			0.04
			Transasia / Erba	16	1.36	9.29	0.13			0.06
			Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
			Any Analyser (Automation / Semi Automation)	70	1.32	7.52	0.10			0.02
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68			2.20
			Transasia / Erba	19	111.05	6.05	6.72			3.08
			Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	122 mg/dL	1.91	0.28
			Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78			0.79
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

