

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name: Govt PHC Neral
 Lab No: 12951
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Back to Search Page

Print

Date of Result Entered: 20/06/2023
 Date of Report Published: 04/07/2023

Sl No	Test Name	Method	Result	Unit	Reference Range	Remarks
1	GLUCOSE III	GOD-POD III	280.51	4.07	11.42	
		Transasia / Erba	283.77	5.80	16.44	
		Robonik (Semi automatic / Fully Automatic)	286.63	5.17	14.82	27.5 mg/dL
		Any Analyser (Automation / Semi Automation)	287.55	4.50	12.95	-0.78
		Biosystems	2.67	6.50	0.17	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	2.50	9.65	0.24	
		Transasia / Erba	2.57	6.64	0.17	2.6 mg/dL
		Robonik (Semi automatic / Fully Automatic)	2.58	6.39	0.16	0.18
		Any Analyser (Automation / Semi Automation)	113.79	4.23	4.81	
		Biosystems	22	5.43	6.22	
3	CHOLESTROL III	CHOD-PAP III	113.73	4.46	5.07	
		Transasia / Erba	114.61	4.46	5.07	
		Robonik (Semi automatic / Fully Automatic)	113.73	4.46	5.07	115 mg/dL
		Any Analyser (Automation / Semi Automation)	114.61	4.05	4.64	0.25
		Biosystems	149	4.05	4.64	0.76

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links

மருத்துவ அறிவுரை
 அரசினர் ஆஸ்பத்திரியில் சுகாதார நிறைவு
 பெறும் - 632 316, திருச்சி மாநகர்