

CALIBRATION CERTIFICATE										
CERTIFICATE NO: SBS/CL/23/12402			3/12402	MECHANICAL			Page No:1 of 1			
Issue Date					05-10-2023					
SRF No & Date					SRF/23/00580-0001 & 03-10-2023					
Receipt Date					03-10-2023					
Calibration Date					03-10-2023					
Calibration Due					02-10-2024					
Customer Name & Address										
GOVERNMENT PRIMARY HEALTH CENTRE,										
NESAL-63230,THIRUVANNAMALAI DISTRICT.										
Details of Device Under Calibration (DUC)										
Description : CENTRIFUGE			JGE	Make & Model :		B.D.INSTRUMENTATION & BDI-152				
Range :			3000 RPM	Sr. N	0	:	2011898			
Resolu			1 RPM	Ident	ification No	:	NA			
DUC C	Condition :	Satisfactor	$\mathbf{v}$	Loca	tion	:	LABORATO	RY		
Environmental Conditions & Standard Operating Procedure Details										
Environmental Details Temperature: 25.5°				°C	C Relative Humidity			52%Rh		
Calibration Procedure No SBS/CP/ML/0			SBS/CP/ML/04	Calibration done at			ONSITE			
Reference Standards Details										
S.No Description Make/				SI No:	SI No: Certificate No		0		Validity	
1 Digital Tachometer LINE S					EIKI / 175-0034V JRPM-CCTR-A&S-2023-00			013	09-06-2024	
CALIBRATION RESULTS										
S.No	DEVICE UNDER CALIBRATION		STANDARD INSTRUMEN		DEVIATION		EXPANDED UNCERTAINTY			
	RPM		RPM		RPM		%			
1	1000		999.0		1.0		4.2			
2	1500		1499.5		0.5		4.2			
3	3000		2999.5		0.5		4.2			
REMARKS   1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.   2. The user should determine the suitability of the instrument for its intended use.   3. The recalibration interval should be determined on the user requirement.   4. The results stated in this certificate relate only to the item calibrated.										
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.										
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.										
Calibrated By Authorised Signatory										
(Calibration Engineer) K.SATHYAMOORTHY										

SUNSHINE BIOMEDICAL SOLUTIONS No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.