



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC Pandukudi
 Lab No: 7467
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Click on the analyte to view Graphical Data

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	283.7 mg/dL	-0.20	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	119.1 mg/dL	1.06	0.24

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.

Links		
Modify Address	Modify Methods	Submit Lab Results
Yearly Summary	Change Password	Home
		View Lab Monthly Summary
		Sign-out

Site Maintained by
Swiftware Solutions