

**NEW LAB MONTHLY SUMMARY**

Lab Name **Govt. PHC, Pallikonda**  
 Lab No **10651**  
 Month **May**  
 Year **2022**  
 Constituent Group **Chemistry III**

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 17/05/2022

Date of Report Published : 08/06/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	146.47	6.38	9.35	128 mg/dL	-1.98	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1771	8.11	8.92	0.72	3 mg/dL	-7.1	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1730	119.85	6.17	7.39	145 mg/dL	3.40	0.36

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.